

EU Type Examination Certificate

This is to certify that:

WRP Asia Pacific Sdn Bhd
Lot 1, Jalan 3
Kawasan Perusahaan
Bandar Baru Salak Tinggi
Sepang
Selangor
43900
Malaysia

Holds Certificate Number:

CE 688314

In respect of:

Powder Free Nitrile Gloves

on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Drs. Dave Hagenaaers, Managing Director

First Issued: 2018-04-26

Latest Issue: 2020-10-27

Effective Date: 2020-10-27

Expiry Date: 2023-04-26

Page: 1 of 5



...making excellence a habit.™

EU Type Examination Certificate

No. CE 688314

Product Specification

Model – Nitrile gloves

Classification: Protective glove for use against chemical and micro-organism hazards.

Description: Powder Free Nitrile Disposable Protective Gloves manufactured from 100% nitrile synthetic rubber latex (Acrylonitrile-butadiene) not containing natural rubber latex. Inner surface of gloves is smooth surface that assists in donning the gloves without using lubricant such as powder on the glove surface. The glove is available as either Sterile or Non-Sterile as indicated.

The main features of this protective glove are:

- Beaded cuff
- Micro-organisms penetration resistance
- Chemical permeation resistance
- Diamond embossed of palm/back of hand (Ambidextrous)

Models in range: Nitrile gloves.

Category: Category III – complex

Applicable General requirements for gloves to EN 420:2003 +A1:2009

Standards: Protective gloves against chemicals & micro-organisms to EN ISO 374-1:2016

Resistance to penetration to EN 374-2:2014

Determination of resistance to permeation by chemicals tested to EN 16523-1:2015

Resistance to penetration by blood borne tested to EN ISO 374-5:2016

Resistance to degradation by chemicals to EN 374-4:2013

Resistance of clothing materials to penetration by blood-borne pathogens to ISO 16604:2014

First Issued: 2018-04-26

Latest Issue: 2020-10-27

Effective Date: 2020-10-27

Expiry Date: 2023-04-26

Page: 2 of 5

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 688314

Product Specification (Continued)

Model – Nitrile disposable gloves (WRP NBR 3549)

Performance: **General requirements for gloves to EN 420:2003+A1:2009**

Dexterity: Level 5

Resistance to penetration to EN 374-2:2014

Pass

**Resistance to chemical permeation to EN ISO 374-1:2016
Test method EN 16523-1:2015**

n-Heptane (J) – level 1
40% Sodium hydroxide (K) – level 6
96% Sulphuric acid (L) – level 0
37% Formaldehyde (T) – level 5
30% Hydrogen peroxide (P) – level 5

Protection against micro-organism risks to EN ISO 374-5:2016

Bacteria and fungi (Test method EN 374-2:2014) Pass
Viruses (Test Method ISO 16604:2004) Pass

Resistance to degradation by chemicals to EN 374-4:2013

n-Heptane (J) 36.7%
40% Sodium hydroxide (K) -9.0%
96% Sulphuric acid (L) 100%
37% Formaldehyde (T) -39.1%
30% Hydrogen peroxide 40.5%

First Issued: 2018-04-26

Latest Issue: 2020-10-27

Effective Date: 2020-10-27

Expiry Date: 2023-04-26

Page: 3 of 5

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 688314

Product Specification (Continued)

Model – Nitrile multipurpose gloves (WRP NBR E DUO)

Performance: General requirements for gloves to EN 420:2003+A1:2009

PH Value: Pass

Resistance to chemical permeation to EN ISO 374-1:2016 Test method EN 16523-1:2015

n-Heptane (J) – level 1
40% Sodium hydroxide (K) – level 6
37% Formaldehyde (T) – level 5
30% Hydrogen peroxide (P) – level 4
40% Hydrofluoric acid – level 1

Resistance to penetration by blood borne tested to EN ISO 374-5:2016
Pass

Resistance to degradation by chemicals to EN 374-4:2013

n-Heptane (J) 38.2%
40% Sodium hydroxide (K) –16.3%
37% Formaldehyde (T) 1.3%
30% Hydrogen peroxide (P) 7.1%

First Issued: 2018-04-26

Latest Issue: 2020-10-27

Effective Date: 2020-10-27

Expiry Date: 2023-04-26

Page: 4 of 5

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 688314

Certificate Administration Details

Technical File Reference: TF/PPE/002

Certificate Amendment Record

Issue date	Comments	BSI Internal report No.
March 2018	First issue.	0086:18:8888423
December 2018	Addition of models Nitrile Multi-Purpose Gloves.	0086:18:9665354
March 2019	Revision to add virus claim.	2797:19:3053773
October 2020	Administration update to revised Module D certificate CE 688305 reference.	2797:20:3317692

Note: The Certificate holder is responsible for keeping the Notified Body advised of changes to any aspect of the overall process used in the manufacture of the product.

Certificate validity

The Certificate holder is responsible for ensuring that the Notified Body is advised of changes to any aspect of the overall process utilised in the manufacture of the product, failure to do so could invalidate the Certificate in respect of product manufactured following the introduction of such changes.

The Conformity to Type Based on Quality Assurance of the Production Process, Annex VIII (Module D), for the product are referenced in BSI issued Certificate number CE 688305.

First Issued: 2018-04-26

Latest Issue: 2020-10-27

Effective Date: 2020-10-27

Expiry Date: 2023-04-26

Page: 5 of 5

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.